

Quotation

No.JJH/PUR/Quotation / 2109 / 2020
Office of the Dean,
Sir J.J.Group of Hospitals, Mumba.
Date 09/01/2020

To,
1) M/s.....
.....
.....
.....

**Sub : Purchase of material for the use of Blood Bank,
Sir J.J. Group of hospitals, Mumbai 400 008.**

Dear Sir,

Sealed quotation is invited to supply the following material for this hospital. Your sealed quotations should be reach this office on or before 16/01/2020

The rate should be quoted for the material specified below. The rate other than the specified will not be considered. Please quote this office reference on the top of the envelope with due date.

Sr.No.	Name of Reagent	Dept.	Qty
1	RPR Card Test for syphilis	रक्तपेढी विभाग	3000 Nos


Dean,
Sir J.J. Group of Hospitals, Mumbai.

N.B.

1. The quotation will be accepted up to 5.00 pm on the working day.f
2. The terms and conditions with delivery date should be mentioned.

Specification for RPR Cards Test for syphilis

- 1) *In vitro* diagnostic reagent for blood bank use only. Not for medicine use.
- 2) The assay should allow for qualitative and semi quantitative determination of Reagin antibodies in serum or plasma for sero-diagnosis of syphilis based on flocculation principle using non treponemal antigens.
- 3) It should have disposable sample/control dispensing pipettes, disposable slides with reaction circles & all essential accessories required.
- 4) It should have Positive & Negative control.
- 5) The assay should be suitable to perform with either serum or plasma.
- 6) The assay should have sensitivity of 85% or more in primary syphilis and a specificity of 93% or more.
- 7) The kit should have long shelf life of 60% or 12 months (whichever is more) at the port of discharge of consignment.
- 8) Storage temperature 2 to 8 degree centigrade.
- 9) The test should be able to yield result within 20 minutes.
- 10) Should be FDA approved for use in blood bank.

Quotation

No.JJH/PUR/Quotation / 2108 /2020
Office of the Dean,
Sir J.J.Group of Hospitals, Mumba.
Date 09/01/2020

To,

- 1) M/s.....
.....
.....
.....

**Sub : Purchase of material for the use of Blood Bank,
Sir J.J. Group of hospitals, Mumbai 400 008.**

Dear Sir,

Sealed quotation is invited to supply the following material for this hospital. Your sealed quotations should be reach this office on or before 17/01/2020

The rate should be quoted for the material specified below. The rate other than the specified will not be considered. Please quote this office reference on the top of the envelope with due date.

Sr.No.	Name of Reagent	Dept.	Qty
1	Microtips (5ul-100ul	रक्तपेढी विभाग	50000 Nos
2	Malaria Test kit (pan/pv/pf)	रक्तपेढी विभाग	3000 Nos
3	Temperature Recording Charts	रक्तपेढी विभाग	500 Nos
4	Anti AB antisera	रक्तपेढी विभाग	100 vials (per vial 10 ml)

pe.
Dean,
Sir J.J. Group of Hospitals, Mumbai.

N.B.

1. The quotation will be accepted up to 5.00 pm on the working day.f
2. The terms and conditions with delivery date should be mentioned.

1. Specification for Anti AB Antisera

1. Antisera must be monoclonal, for use in blood bank.
2. Should be of high quality with a shelf life of at-least one year of use and should be received in cold chain .
3. Should contain a preservative to minimize contamination.
4. Storage temperature: 2-8°C.
5. Reagent must be clear.
6. Clear-cut reaction with RBC of A and B group.
7. Titer 1:256 in reaction with B cells and A cells.
8. Avidity 3-4 sec.

2. Specification for Microtips

For use with single channel micropipette in blood bank

1. Should be made of plastic.
2. Firmly fits onto micropipette.
3. Volume 5- 200 μ l.
4. Non sterile.

3. Specification for Temperature Recording Charts

For Blood bank Refrigerator (400 nos)

1. Should be round, for use in blood bank refrigerator for recording temperature.
2. Should be recordable in **Ink**.
3. Should be of high quality for long term storage.
4. Should have clearly legible markings of temperature and days.
5. Should be suitable for use in Terumo Penpol blood bank refrigerator.

Specification for Temperature Recording Charts For Platelet Agitator (100 nos)

1. Should be round, for use in blood bank for recording temperature of platelet agitator.
2. Should be recordable , in temperature (**thermo**) sensitive type.
3. Should be of high quality for long term storage.
4. Should have clearly legible markings of temperature and days.

4. Specification for Malaria Test kits (pan/pv/pf)

- 1) Application : Malaria rapid card test should be visual , rapid and sensitive immunoassay for the qualitative diagnosis of all four Plasmodium Species (pan/pv/pf) in human whole blood. The method should use anti – pan to detect Malaria antigen.
- 2) The test should use anti – pan specific PLDH.
- 3) Sensitivity : 99.6%
- 4) Specificity : 100%
- 5) Storage /Shelf Life : Storage – 4 -30°C/Stability - 2 year (24 Months)
- 6) Should have Control band & should include all essential accessories.


Dean,

Sir J.J. Group of Hospitals, Mumbai.

Quotation

No.JJH/PUR/Quotation/2107 12020
Office of the Dean,
Sir J.J.Group of Hospitals, Mumbai.
Date :09/01/2020

To,

- 1) M/s.....
.....
.....
.....

**Sub : Purchase of material for the use of Blood Bank,
Sir J.J. Group of hospitals, Mumbai 400 008.**

Dear Sir,

Sealed quotation is invited to supply the following material for this hospital. Your sealed quotations should be reach this office on or before 16/01/2020

The rate should be quoted for the material specified below. The rate other than the specified will not be considered. Please quote this office reference on the top of the envelope with due date.

Sr.No.	Name of Reagent	Dept.	Qty
1	Blood Bags Labels	रक्तपेढी विभाग	4600Nos
2	Blood Bags Serial No. Stickers	रक्तपेढी विभाग	56000 Nos

Specification for CPD Bags Stickers

1. Should be made of suitable firm moisture resistant material for sticking on CPD blood bags.
2. Material should have good quality clearly readable printing in black and red.
3. Should have colour coding for blood groups as per international norms
4. Should have good quality adhesive to ensure that label sticks to blood bag firmly for the entire shelf life.
5. Manual writing of details eg date of collection etc, should be possible with marker pen.


Dean,

Sir J.J. Group of Hospitals, Mumbai.

N.B.

1. The quotation will be accepted up to 5.00 pm on the working day.
2. The terms and conditions with delivery date should be mentioned.